

Analysis of Aroclor 1260 on a 50m HT8

KEYWORDS:

Aroclor 1260


PCB

Components

PCB Congeners are labelled using IUP nomenclature

Part No.: 054676
Phase: HT8, 0.25µm
Column: 50m x 0.22mm I.D.
Initial Temp.: 80°C, 2 min.
Rate 1: 30°C/min
Temp. 2 : 170°C
Rate 2: 3°C/min
Final Temp.: Split, 300°C
Detector: ECD, 330°C
Carrier Gas: He, 40psi

Analysis of Aroclor 1260



visit www.sge.com or contact our technical Customer Support Team on: techsupport@sge.com

SGE International Pty. Ltd. email: support@sge.com	SGE Incorporated (USA) email: usa@sge.com	SGE (France) Sarl email: france@sge.com	SGE Deutschland GmbH email: germany@sge.com	SGE Japan Inc. email: sgejpn@itjit.ne.jp
SGE Australia Pty. Ltd. email: sales@sge.com	SGE Europe Ltd. (UK) email: uk@sge.com	SGE (Italia) Srl. email: sge.italia@tin.it	SGE China Service Centre email: sge@bj.col.com.cn	SGE India email: sgeindia@bom4.vsnl.net.in